

CALIBRATION CERTIFICATE


Company:	I Waste	Certificate No:	252825
Address:	Unit 4 Ducks Nest Farm	Postcode:	RG2 9PJ
	Eversley Road	Contact:	Anna Webb
	Arborfield Cross. Reading	Telephone:	0330 120 0250

Load And Linearity Test

Applied Load:	Kg	As Found Reading	Deviation	Definitive Reading	Deviation
0		0.0	0.0	0.0	0.0
50		50.0	0.0	50.0	0.0
100		100.0	0.0	100.0	0.0
200		199.5	-0.5	199.5	-0.5
300		299.5	-0.5	299.5	-0.5
400		399.0	-1.0	399.0	-1.0

Eccentric Test

Applied Load:	Kg	As Found Reading	Deviation	Definitive Reading	Deviation
1	200	199.5	-0.5	199.5	-0.5
2	200	199.5	-0.5	199.5	-0.5
3	200	199.5	-0.5	199.5	-0.5
4	200	199.5	-0.5	199.5	-0.5



Equipment Details

Manufacturer:	Monster Shop	Model:	
Serial Number:	001	Location:	Yard
Capacity:	3000kg	Division Size: e	0.5kg
Range:	400kg	Tolerances:	<500e 1e 501 - 2000e 2e >2000e 3e

Repeatability Test

Test One

Test Two

Test Three

Applied Load:	400kg		
Definitive Reading:	399.0	399.0	399.0
Test Weight Refs:	Norfolk C.C UKAS Lab 0260, Certs: UP2146, UP2281/82, UN2540, UR0249		

Engineers Report

Date of Test:	17 October 2025
Next Calibration Due:	October 2026
Comments:	
Tested By:	P. Thurgood
Signed:	

These tests comply with the code of practice of the UK Weighing Federation



Scan QR Code
For our Latest
ISO 9001:2015 Certificate

Solent Scale Services
Unit 9 Quarry Lane
Chichester
West Sussex PO19 8QA

Tel: 01243 697372
Email: admin@solentscales.co.uk
Registered in England No. 4031406
VAT Registration No. 760233849